


## Editorial



# Sapien 3 versus Sapien XT Balloon-expanding Valve for Transcatheter Aortic Valve Replacement: Improved Immediate and Late Outcomes at the Expense of Smaller Effective Orifice Area

Young-Guk Ko , MD

Division of Cardiology, Severance Cardiovascular Hospital, Yonsei University College of Medicine, Seoul, Korea

## OPEN ACCESS

► See the article “Comparing the Procedural and Clinical Outcomes of Sapien XT and Sapien 3 Valves in Transcatheter Aortic Valve Replacement in Korean Patients” in volume 50 on page 907.

**Received:** Aug 6, 2020

**Accepted:** Aug 18, 2020

### Correspondence to

Young-Guk Ko, MD

Division of Cardiology, Severance Cardiovascular Hospital, Yonsei University College of Medicine, 50-1, Yonsei-ro, Seodaemun-gu, Seoul 03722, Korea.  
E-mail: ygko@yuhs.ac

Copyright © 2020. The Korean Society of Cardiology

This is an Open Access article distributed under the terms of the Creative Commons Attribution Non-Commercial License (<https://creativecommons.org/licenses/by-nc/4.0>) which permits unrestricted noncommercial use, distribution, and reproduction in any medium, provided the original work is properly cited.

### ORCID iDs

Young-Guk Ko   
<https://orcid.org/0000-0001-7748-5788>

### Conflict of Interest

The author has no financial conflicts of interest.

The contents of the report are the author's own views and do not necessarily reflect the views of the *Korean Circulation Journal*.

Transcatheter aortic valve replacement (TAVR) has successfully demonstrated its efficacy and safety as an alternative therapy to surgical aortic valve replacement not only in high or intermediate risk patients but also in low risk patients with severe aortic valve stenosis.<sup>1,2)</sup> However, TAVR is associated with morbidities such as paravalvular leak (PVL), conduction abnormalities, and vascular complications.<sup>2,3)</sup> Sapien valve system (Edwards Lifesciences, Irvine, CA, USA) was the first commercially available TAVR prosthesis for clinical use.<sup>4)</sup> Sapien 3 is the latest generation valve currently available in Korea. Whereas the previous generation Sapien XT used an introducer sheath of 16–18F diameter, Sapien 3 requires a lower profile sheath of 14–16F. In addition, Sapien 3 added an outer sealing skirt made of polyethylene terephthalate to minimize paravalvular leak.

In this issue of *Korean Circulation Journal*, Kook et al.<sup>5)</sup> reported outcomes of Sapien 3 in TAVR in Korean patients compared with Sapien XT. In their study, Sapien 3 showed a higher device success rate and a lower incidence of moderate or severe PVL than Sapien XT. The 1-year cardiovascular mortality rate was also lower with Sapien 3. This study has limitations including retrospective study design, small sample size, inadequate adjustment of confounding factors, especially learning curve of procedure skills and difference experience levels among participating centers. Despite these limitations, the findings of the study are generally consistent with the previous reports.<sup>6,7)</sup> A recent meta-analysis study demonstrated that Sapien 3 achieved higher rates of device success and lower rates of moderate or severe PVL, major vascular complications, major bleeding, stroke, and early mortality than Sapien XT.<sup>6)</sup> Thus, the lower delivery system profile and the outer skirt of the Sapien 3 appear to successfully reduce procedure-related complications and significant PVL and to improve the device success. However, permanent pacemaker implantation rate was higher after implantation of Sapien 3 rather than Sapien XT. The outer skirt of Sapien 3 may be helpful to reduce PVL, however it seems to induce more atrioventricular conduction abnormality as a trade-off. However, in the present study, the pacemaker implantation rate was very low for both Sapien XT and Sapien 3 (2.1% vs. 1.1%,  $p=0.621$ ). Whether operator techniques or patient ethnicity might contribute to the low permanent pacemaker implantation rate remains unknown.

Kook et al.<sup>5)</sup> reported in this study that the effective orifice area of the implanted valve was smaller with Sapien 3 than Sapien XT and suggested that the outer skirt of Sapien 3 might cause a smaller effective orifice area. Similar findings have been also observed in the other studies. Abdelghani et al.<sup>8)</sup> speculated lesser valve oversizing as another explanation for the smaller effective orifice area (EOA) with Sapien 3. They described that there was a significant decline of percentage oversizing of the balloon expanding valves from 11.4% with Sapien XT to 3.7% with Sapien 3 despite similar manufacturer's sizing recommendation. The authors explained that the data on annulus rupture by excessive oversizing in cases of Sapien XT might have contributed in part to more conservative approach in valve sizing since Sapien 3 is capable to reduce PVL with the outer sealing skirt. Thus, from the safety point of view, lesser oversizing of the Sapien 3 appears to be appropriate. However, smaller EOA is associated with a higher residual transvalvular pressure gradient and a higher incidence of patient-prosthesis mismatch (PPM) especially in cases of small valve sizes. Theron et al.<sup>9)</sup> demonstrated that the risk of PPM is increased with 23 mm Sapien 3 in comparison with 23 mm Sapien XT. PPM has been shown to be associated with an increase in all-cause and cardiac-related mortality over long-term follow-up.<sup>10)</sup> Therefore, Sapien valve system needs to find a solution to increase the EOA and thereby to reduce the risk of PPM especially in cases of small valve sizes.

## REFERENCES

1. Al-Azizi K, Hamandi M, Mack M. Clinical trials of transcatheter aortic valve replacement. *Heart* 2019;105:s6-9.  
[PUBMED](#) | [CROSSREF](#)
2. Kim C, Hong MK. Aortic stenosis and transcatheter aortic valve implantation: current status and future directions in Korea. *Korean Circ J* 2019;49:283-97.  
[PUBMED](#) | [CROSSREF](#)
3. Yu CW, Kim WJ, Ahn JM, et al. Trends and outcomes of transcatheter aortic valve implantation (TAVI) in Korea: the results of the first cohort of Korean TAVI registry. *Korean Circ J* 2018;48:382-94.  
[PUBMED](#) | [CROSSREF](#)
4. Ribeiro HB, Urena M, Allende R, Amat-Santos IJ, Rodés-Cabau J. Balloon-expandable prostheses for transcatheter aortic valve replacement. *Prog Cardiovasc Dis* 2014;56:583-95.  
[PUBMED](#) | [CROSSREF](#)
5. Kook H, Jang DH, Yang KS, et al. Comparing the procedural and clinical outcomes of Sapien XT and Sapien 3 valves in transcatheter aortic valve replacement in Korean patients. *Korean Circ J* 2020;50:907-22.  
[PUBMED](#) | [CROSSREF](#)
6. Tummala R, Banerjee K, Sankaramangalam K, et al. Clinical and procedural outcomes with the SAPIEN 3 versus the SAPIEN XT prosthetic valves in transcatheter aortic valve replacement: A systematic review and meta-analysis. *Catheter Cardiovasc Interv* 2018;92:E149-58.  
[PUBMED](#) | [CROSSREF](#)
7. Ando T, Briasoulis A, Holmes AA, Taub CC, Takagi H, Afonso L. Sapien 3 versus Sapien XT prosthetic valves in transcatheter aortic valve implantation: A meta-analysis. *Int J Cardiol* 2016;220:472-8.  
[PUBMED](#) | [CROSSREF](#)
8. Abdelghani M, Mankerious N, Allali A, et al. Bioprosthetic valve performance after transcatheter aortic valve replacement with self-expanding versus balloon-expandable valves in large versus small aortic valve annuli: insights from the CHOICE trial and the CHOICE-Extend registry. *JACC Cardiovasc Interv* 2018;11:2507-18.  
[PUBMED](#) | [CROSSREF](#)
9. Theron A, Pinto J, Grisoli D, et al. Patient-prosthesis mismatch in new generation trans-catheter heart valves: a propensity score analysis. *Eur Heart J Cardiovasc Imaging* 2018;19:225-33.  
[PUBMED](#) | [CROSSREF](#)
10. Head SJ, Mokhles MM, Osnabrugge RL, et al. The impact of prosthesis-patient mismatch on long-term survival after aortic valve replacement: a systematic review and meta-analysis of 34 observational studies comprising 27 186 patients with 133 141 patient-years. *Eur Heart J* 2012;33:1518-29.  
[PUBMED](#) | [CROSSREF](#)